

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
Name <u>DARYL RICE</u>	ID# <u>2088</u> Date <u>03/10/2012</u>
Agency <u>SITKA POLICE DEPARTMENT</u> Phone # <u>747-3245</u>	
Instrument Location <u>304 LAKE STREET</u>	
Alco S/N <u>X172969</u> Target Value <u>.077</u> High Pressure <u>550</u>	
Alco Test Values 1st Alco <u>.079</u> 2nd Alco <u>.078</u>	
Signature <u>Daryl Rice</u>	

(OVER)

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 5 day of April, 2012.

Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office

(Notary Seal Stamp)



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130176

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130176

MARCH 10, 2012

OPERATOR'S NAME:

RICE DARYL DEAN

OPERATOR'S NUMBER: 2088

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: SIT1

CASE/REPORT: VOC

TEST TYPE: U

ALCO TARGET VALUE: .077

ALCO S/N: X172969

### — BREATH ANALYSIS —

.077 ADJUSTED FOR 29.51 in

ALCO TARGET .075 04:19

BLANK TEST .000 04:20

INTERNAL STANDARD VERIFIED 04:21

ALCO TV 29.51 in .079 04:21

BLANK TEST .000 04:22

SUBJECT SAMPLE .000 04:22

BLANK TEST .000 04:23

ALCO TV 29.51 in .078 04:23

BLANK TEST .000 04:24

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130176

MARCH 10, 2012

TIME 04:24

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.51 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

!@#%&'()\*+,-./0123456789:;<=>?@BCDEFG  
HIJKL/MNOPQRSTUVWXYZ\]^\_`'abdefghijklmno  
pqrstuvwxyz{|}~